


UREA SOLUTION

alternative for surface chip regeneration

Dynamic Biosensors GmbH & Inc.
SOL-REG-UR-12-1 v2.1



Product Description

Order Number: **SOL-REG-UR-12-1**

Table 1. Contents and Storage Information

Material	Composition	Amount	Storage
Urea solution	7 M Urea; 0.2 µm sterile filtered	12 mL	2-8°C

For research use only.

This product has a limited shelf life, please see expiry date on label.

Please be aware that this solution is not suitable for a complete regeneration of the chip as only proteins will be removed. The complementary nanolever will stay on the chip.

Useful Order Numbers

Table 2. Order Numbers

Product Name	Comment	Order Number
helix[®] Adapter Chip	Chip with 2 detection spots	ADP-48-2-0
10x Passivation solution	For passivation of chip surface	SOL-PAS-1-5
Regeneration solution	For regeneration of chip surface	SOL-REG-1-5

Contact

Dynamic Biosensors GmbH
Perchtinger Str. 8/10
81379 Munich
Germany

Dynamic Biosensors, Inc.
300 Trade Center, Suite 1400
Woburn, MA 01801
USA

Order Information order@dynamic-biosensors.com
Technical Support support@dynamic-biosensors.com

www.dynamic-biosensors.com

Instruments and chips are engineered and manufactured in Germany.
©2024 Dynamic Biosensors GmbH | Dynamic Biosensors, Inc. All rights reserved.