5x Buffer B pH 7.2

After dilution this buffer is intended for use as the **pro**FIRE[®] Buffer B.

Product Description

Order Number BU-P-1000-5

Material	Amount	Storage	Comment
5x Buffer B pH 7.2 (250 mM Na ₂ HPO ₄ /NaH ₂ PO ₄ and 5 M NaCl; 0.2 μm sterile filtered)	50 mL	r.t.	

For *in vitro* use only.

This product has a limited stability, please see expiry date on label. At temperatures lower than 18°C precipitation may occur.

Preparation

Dilute 5x Buffer B pH 7.2 with deionized, 0.2 µm filtered water by mixing the complete solution (50 mL) with 200 mL ultrapure water. After dilution Buffer B is ready for use (50 mM Na₂HPO₄/NaH₂PO₄ pH 7.2 and 1 M NaCl).

The diluted buffer should be stored at 2-8°C.




Contact

Dynamic Biosensors GmbH

Perchtinger Str. 8/10 81379 Munich Germany Phone: +49 89 89 74 544 0 Dynamic Biosensors Inc. 300 Trade Center, Suite 1400 Woburn, MA 01801 USA Phone: +1 781 404 6126

Order Information	Phone: +49 89 89 74 544 0
	Email: order@dynamic-biosensors.com
Technical Support	Phone: +49 89 89 74 544 66
	Email: support@dynamic-biosensors.com



switchSENSE[®] is a proprietary measurement technology by Dynamic Biosensors GmbH. Instruments and biochips are engineered and manufactured in Germany.

©2023 Dynamic Biosensors GmbH | Dynamic Biosensors Inc. All rights reserved.