## 10x Buffer P5 pH 7.3

After dilution this buffer is intended for use as a running buffer with the DRX family.

## **Product Description**

Order Number BU-P-5-10

Material	Amount	Storage	Comment
10x Buffer P5 pH 7.3 (50 mM NaH <sub>2</sub> PO <sub>4</sub> /Na <sub>2</sub> HPO <sub>4</sub> ; 0.2 µm sterile filtered)	50 mL	2-8°C	

For in vitro use only.

This product has a limited stability, please see expiry date on label.

## **Preparation**

Dilute 10x Buffer P5 pH 7.3 with deionized, 0.2  $\mu$ m filtered water by mixing the complete solution (50 mL) with 450 mL ultrapure water. After dilution P5 Buffer is ready for use (5 mM NaH<sub>2</sub>PO<sub>4</sub>/Na<sub>2</sub>HPO<sub>4</sub>, pH 7.3).

The diluted buffer should be stored at 2-8°C.



## **Contact**

**Dynamic Biosensors GmbH** 

Perchtinger Str. 8/10 81379 Munich Germany

Phone: +49 89 89 74 544 0

**Dynamic Biosensors Inc.** 

300 Trade Center, Suite 1400

Woburn, MA 01801

USA


Phone: +1 781 404 6126

**Order Information** Phone: +49 89 89 74 544 0

Email: order@dynamic-biosensors.com

**Technical Support** Phone: +49 89 89 74 544 66

Email: support@dynamic-biosensors.com



**switch**SENSE® is a proprietary measurement technology by Dynamic Biosensors GmbH. Instruments and biochips are engineered and manufactured in Germany.

©2023 Dynamic Biosensors GmbH | Dynamic Biosensors Inc. All rights reserved.