

10x Buffer A pH 7.2

After dilution this buffer is intended for use as the **proFIRE[®]** Buffer A.

Product Description

Order Number **BU-P-150-10**

Material	Amount	Storage	Comment
10x Buffer A pH 7.2 (500 mM Na ₂ HPO ₄ /NaH ₂ PO ₄ and 1.5 M NaCl; 0.2 µm sterile filtered)	50 mL	r.t.	

For *in vitro* use only.

This product has a limited stability, please see expiry date on label.
At temperatures lower than 18°C precipitation may occur.

Preparation

Dilute 10x Buffer A pH 7.2 with deionized, 0.2 µm filtered water by mixing the complete solution (50 mL) with 450 mL ultrapure water. After dilution Buffer A is ready for use (50 mM Na₂HPO₄/NaH₂PO₄ pH 7.2 and 150 mM NaCl).

The diluted buffer should be stored at 2-8°C.

Contact

Dynamic Biosensors GmbH

Perchtinger Str. 8/10
81379 Munich
Germany

Phone: +49 89 89 74 544 0

Dynamic Biosensors Inc.

300 Trade Center, Suite 1400
Woburn, MA 01801
USA

Phone: +1 781 404 6126

Order Information


Phone: +49 89 89 74 544 0

Email: order@dynamic-biosensors.com

Technical Support

Phone: +49 89 89 74 544 66

Email: support@dynamic-biosensors.com



switchSENSE® is a proprietary measurement technology by Dynamic Biosensors GmbH. Instruments and biochips are engineered and manufactured in Germany.

©2023 Dynamic Biosensors GmbH | Dynamic Biosensors Inc. All rights reserved.