

## Digoxigenin kit for 96mer

This kit is intended for use as an immobilization method for anti-Digoxigenin antibodies or Fab fragments with a pM affinity.

It can be used with the following **switch**SENSE® *Multi-purpose chips* for kinetic experiments:

MPC-96-1-Y1-X, MPC-96-1-R1-X, MPC-96-2-Y1-X, MPC-96-2-R1-X, MPC2-96-2-G1R1-X and MP3-96-2-G1R1-X.

### Product Description

Order Number **CK-DG-1-B96** (nanolever sequence B96)

Material	Cap	Amount	Storage	Comment
cNL-B96-Digoxigenin (Digoxigenin conjugated to cNL-B96, 400 nM in PE40 <sup>1</sup> )	green	20 x 20 µL	-20°C	
cNL-A96 (400 nM in TE40 <sup>2</sup> )	yellow	4 x 100 µL	-20°C	

For *in vitro* use only.

The kit contains reagents sufficient for 20 regenerations.

This product has a limited stability, please see expiry date on label.

### Preparation/Immobilization

For one immobilization: Mix 20 µL of cNL-B96-Digoxigenin (400 nM) with 20 µL cNL-A96 (400 nM).

In switchBUILD choose assay type kinetics and select Sandwich Format for immobilization. Set mixture as ligand and place it according to autosampler layout.

<sup>1</sup> 10 mM Na<sub>2</sub>HPO<sub>4</sub>/NaH<sub>2</sub>PO<sub>4</sub>, 40 mM NaCl, 0.05 % Tween20, 50 µM EDTA, 50 µM EGTA

<sup>2</sup> 10 mM Tris, 40 mM NaCl, 0.05 % Tween20, 50 µM EDTA, 50 µM EGTA

## Contact

### **Dynamic Biosensors GmbH**

Perchtinger Str. 8/10  
81379 Munich  
Germany

Phone: +49 89 89 74 544 0

### **Dynamic Biosensors Inc.**

300 Trade Center, Suite 1400  
Woburn, MA 01801  
USA

Phone: +1 781 404 6126

### **Order Information**


Phone: +49 89 89 74 544 0

Email: [order@dynamic-biosensors.com](mailto:order@dynamic-biosensors.com)

### **Technical Support**

Phone: +49 89 89 74 544 66

Email: [support@dynamic-biosensors.com](mailto:support@dynamic-biosensors.com)



**switchSENSE®** is a proprietary measurement technology by Dynamic Biosensors GmbH. Instruments and biochips are engineered and manufactured in Germany.

©2023 Dynamic Biosensors GmbH | Dynamic Biosensors Inc. All rights reserved.